



## King's Research Portal

DOI:

[10.1186/s12968-018-0502-7](https://doi.org/10.1186/s12968-018-0502-7)

*Document Version*

Peer reviewed version

[Link to publication record in King's Research Portal](#)

*Citation for published version (APA):*

Nazir, M. S., Neji, R., Speier, P., Reid, F. D. A., Staeb, D., Schmidt, M., ... Roujol, S. (2018). Simultaneous Multi Slice (SMS) SSFP first-pass myocardial perfusion MRI with iterative reconstruction at 1.5T. *Journal of cardiovascular magnetic resonance : official journal of the Society for Cardiovascular Magnetic Resonance*, 20(1). <https://doi.org/10.1186/s12968-018-0502-7>

### Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

### General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

### Take down policy

If you believe that this document breaches copyright please contact [librarypure@kcl.ac.uk](mailto:librarypure@kcl.ac.uk) providing details, and we will remove access to the work immediately and investigate your claim.

# **Simultaneous Multi Slice (SMS) SSFP first-pass myocardial perfusion MRI with iterative reconstruction at 1.5T**

Muhummad Sohaib Nazir<sup>1</sup>, Radhouene Neji<sup>1,2</sup>, Peter Speier<sup>3</sup>, Fiona Reid<sup>4</sup>,  
Daniel Stäb<sup>5</sup>, Michaela Schmidt<sup>3</sup>, Christoph Forman<sup>3</sup>, Reza Razavi<sup>1</sup>, Sven Plein<sup>1</sup>,  
Tevfik F. Ismail<sup>1</sup>, Amedeo Chiribiri<sup>1</sup>, Sébastien Roujol<sup>1</sup>

<sup>1</sup>School of Biomedical Engineering and Imaging Sciences, King's College London, St  
Thomas' Hospital, Westminster Bridge Road, London, SE1 7EH, United Kingdom

<sup>2</sup>MR Research Collaborations, Siemens Healthcare Limited, Frimley, United  
Kingdom

<sup>3</sup>Siemens Healthcare, Erlangen, Germany

<sup>4</sup>Division of Health and Social Care Research, King's College London, London,  
United Kingdom

<sup>5</sup>Siemens Healthcare Pty Ltd, Melbourne, Australia

1 **Author email addresses**

2	Muhummad Sohaib Nazir	sohaib.nazir@gmail.com
3	Radhouene Neji	radhouene.neji@kcl.ac.uk
4	Peter Speier	peter.speier@siemens-healthineers.com
5	Fiona Reid	fiona.reid@kcl.ac.uk
6	Daniel Stäb	daniel.staeb@siemens-healthineers.com
7	Michaela Schmidt	michaelia.schmidt@siemens-healthineers.com
8	Christoph Forman	christoph.forman@siemens-healthineers.com
9	Reza Razavi	reza.razavi@kcl.ac.uk
10	Sven Plein	s.plein@leeds.ac.uk
11	Tevfik F Ismail	tevfik.ismail@kcl.ac.uk
12	Amedeo Chiribiri	amedeo.chiribiri@kcl.ac.uk
13	Sébastien Roujol	sebastien.roujol@kcl.ac.uk

1    **Corresponding author**  
2    Dr Sohaib Nazir  
3    School of Biomedical Engineering and Imaging Sciences  
4    King's College London  
5    3<sup>rd</sup> Floor Lambeth Wing  
6    St Thomas' Hospital  
7    Westminster Bridge Road  
8    London SW1 7EH  
9    United Kingdom  
10   sohaib.nazir@gmail.com

## **Abstract**

### Background

Simultaneous-Multi-Slice (SMS) perfusion imaging has the potential to acquire multiple slices, increasing myocardial coverage without sacrificing in-plane spatial resolution. To maximise signal-to-noise ratio (SNR), SMS can be combined with a balanced steady state free precession (bSSFP) readout. Furthermore, application of gradient-controlled local Larmor adjustment (GC-LOLA) can ensure robustness against off-resonance artifacts and SNR loss can be mitigated by applying iterative reconstruction with spatial and temporal regularisation. The objective of this study was to compare myocardial perfusion imaging using SMS bSSFP imaging with GC-LOLA and iterative reconstruction to 3 slice bSSFP.

### Methods

Two contrast-enhanced rest perfusion sequences were acquired in random order in 8 patients: 6-slice SMS bSSFP and 3 slice bSSFP. All images were reconstructed with TGRAPPA. SMS images were also reconstructed using a non-linear iterative reconstruction with L1 regularisation in wavelet space (SMS-iter) with 7 different combinations for spatial ( $\lambda_\sigma$ ) and temporal ( $\lambda_\tau$ ) regularisation parameters. Qualitative ratings of overall image quality (0=poor image quality, 1=major artifact, 2=minor artifact, 3=excellent), perceived SNR (0=poor SNR, 1=major noise, 2=minor noise, 3=high SNR), frequency of sequence related artifacts and patient related artifacts were undertaken. Quantitative analysis of contrast ratio (CR) and percentage of dark rim artifact (DRA) was performed.

### Results

Among all SMS-iter reconstructions, SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) was identified as the optimal reconstruction with the highest overall image quality, least sequence related artifact and higher perceived SNR. SMS-iter 6 had superior overall image quality ( $2.50 \pm 0.53$  vs  $1.50 \pm 0.53$ ,  $p=0.005$ ) and perceived SNR ( $2.25 \pm 0.46$  vs  $0.75 \pm 0.46$ ,  $p=0.010$ ) compared to 3 slice bSSFP. There were no significant differences in sequence related artifact, CR ( $3.62 \pm 0.39$  vs  $3.66 \pm 0.65$ ,  $p=0.88$ ) or percentage of DRA ( $5.25 \pm 6.56$  vs  $4.25 \pm 4.30$ ,  $p=0.64$ ) with SMS-iter 6 compared to 3 slice bSSFP.

### Conclusions

SMS bSSFP with GC-LOLA and iterative reconstruction improved image quality compared to a 3 slice bSSFP with doubled spatial coverage and preserved in-plane spatial resolution. Future evaluation in patients with coronary artery disease is warranted.

### **Keywords**

Cardiovascular magnetic resonance; myocardial perfusion imaging; simultaneous multi-slice; image acceleration; iterative reconstruction

## 1    **Background**

2    First-pass contrast enhanced myocardial perfusion cardiovascular magnetic  
3    resonance (CMR) is recommended in international guidelines for ischaemia testing  
4    in patients with intermediate risk of coronary artery disease (CAD) [1, 2]. A recent  
5    meta-analysis demonstrated a sensitivity of 89% and specificity of 76% for the  
6    detection of angiographically defined CAD [3].

7    Various pulse sequences are used in clinical practice and guidelines recommend  
8    acquiring at least 3 short axis slices with an in-plane resolution of  $<3 \times 3 \text{ mm}$  [4]. The  
9    sequences used typically employ ECG triggering, saturation pre-pulses, and three to  
10   four sequentially acquired 2D slices distributed over a single heartbeat. Alternatively,  
11   3D techniques have been proposed to achieve whole heart coverage [5, 6] but are  
12   usually associated with reduced in-plane spatial resolution, longer imaging readout  
13   and are more susceptible to respiratory motion [6]. There is considerable debate as  
14   to whether in-plane spatial resolution or spatial coverage are more important for  
15   clinical practice [5].

16   Simultaneous multi-slice (SMS) imaging is an alternative data acquisition strategy [7-  
17   11] with potential to increase spatial coverage without sacrificing in-plane spatial  
18   resolution. Using multiband radiofrequency (RF) pulses, separate anatomical slices  
19   are excited simultaneously. By means of Controlled Aliasing in Parallel Imaging  
20   Results in Higher Acceleration (CAIPIRINHA) [11], the simultaneously excited slices  
21   are shifted with respect to each other in image space, which facilitates their  
22   separation using parallel imaging techniques [12-14]. Hence, multiple slice  
23   acquisitions can be performed in the same duration as a single slice acquisition.

SMS bSSFP can be achieved using linear slice specific RF phase cycles with different RF phase increments between succeeding RF pulses in the simultaneously excited slices [15]. Using different phase cycles in the individual slices renders the frequency response slice specific and results in an increased sensitivity to field inhomogeneities. Gradient controlled local Larmor adjustment (GC-LOLA) can be used to restore the frequency response and thus mitigate the effects of field inhomogeneities by unbalancing the gradients along the slice select direction [16]. Standard parallel imaging techniques such as with GRAPPA [12] and SENSE [13] are associated with SNR degradation in the presence of noise and high undersampling factors. The use of prior information in the form of additional regularisation constraints in the reconstruction can be used to improve the quality and SNR of the reconstructed images [17, 18]. Regularisation can be achieved by assuming the sparsity of the data in a given transform domain [19, 20], as developed in the compressed sensing theory [21]. The reconstruction problem is in this case often formulated as an inverse problem which can be solved using an iterative reconstruction process.

The objective of this study was to determine the feasibility of first-pass myocardial perfusion CMR using SMS with a bSSFP sequence, GC-LOLA and iterative reconstruction and compare to a 3 slice bSSFP sequence in patients.

## **Methods**

### Study population

Patients (n=8) who were referred for a clinically indicated contrast enhanced non-stress CMR scan were prospectively recruited to undergo two additional rest



myocardial perfusion scans. The clinical indication for the scan was for assessment of possible cardiomyopathy (n=5) and assessment of left ventricular volumes and function (n=3). Exclusion criteria were contraindication to gadolinium contrast agent or MRI (non MRI safe metallic implant). The study was approved by the National Research Ethics Service (15/NS/0030) with written informed consent obtained from all patients for inclusion in the study and additional imaging during their clinical CMR scan.

#### Perfusion protocol

Prior to imaging, patients were coached for breath holding and instructed to breath hold during first-pass of contrast. Two rest perfusions scans were acquired in each patient for 3 slice bSSFP and 6-slice SMS, separated by a minimum of 15 minutes to allow for contrast washout. The sequence order was alternated in successive patients in order to negate the effect of higher baseline signal following contrast administration of the first perfusion sequence.

Contrast was administered using a dual bolus technique as previously described [22], with 0.0075 + 0.075mmol/kg of body weight gadolinium (gadobutrol, Gadovist, Bayer, Germany). The prebolus and main bolus contrast were separated by a 25 second delay and injected at a rate of 4mL/s followed by a 25ml flush of normal saline. Each injection was performed by a power injector (Spectris Solaris® EP, MEDRAD, INC., USA).

Slice locations were planned using the systolic phase of the 4 and 2 chamber cine images, and the 3 chamber cine image to ensure the basal slice did not encroach on the left ventricular outflow tract (LVOT). For the 3 slice bSSFP approach, the '3 of 5'

rule was used in order to establish basal, mid and apical slices [23]. This involves planning of 5 equidistant slices from proximal basal left ventricle (LV) from the mitral valve annulus to outer boundary of the LV apex in systole, after which the number of slices is adjusted to 3. For the SMS approach, a '6 of 10' rule was employed to obtain 6 slice locations from the basal LV to the apex. This involved planning 10 equidistant slices from the basal LV to apex in the 4, 3 and 3 chamber cine view in systole and then switching to 6 slices in that orientation.

#### Data acquisition and image reconstruction

Imaging was undertaken at 1.5T (MAGNETOM Aera, Siemens Healthcare, Erlangen, Germany). Sequence parameters were matched between the two sequences: FOV 332 x 332mm, acquired voxel size 1.9 x 1.9mm, slice thickness 10mm, flip angle 50°, bandwidth 1093 Hz/px, in-plane acceleration 2.5, inversion time (TI) 95ms [bSSFP] 130ms [SMS], repetition time (TR) 2.5ms [bSSFP] 2.9ms [SMS], echo time (TE) 1.04ms [bSSFP] 1.24ms [SMS]. Standard bSSFP imaging was acquired with 3 short axis slices. SMS images were acquired with a prototype of a SMS-bSSFP sequence that implements the GC-LOLA technique with a slice acceleration factor of 2 to acquire 6 short axis slices per heartbeat. Two different RF phase cycles were used for the SMS sequences using two different phase increments of  $-\pi/2$  for slice 1 (i.e. 0,  $3\pi/2$ ,  $\pi$ ,  $\pi/2$ , 0, ...) and  $\pi/2$  Slice 2 (i.e. 0,  $\pi/2$ ,  $\pi$ ,  $3\pi/2$ , 0, ...). GC-LOLA was used to compensate for the slice specific shifts of the bSSFP frequency response induced by these RF phase cycles as previously described [16]. SMS data were reconstructed using a prototype of a non-linear iterative reconstruction with L1 regularisation in wavelet space (referred to as SMS-iter) [19, 21, 24],

implemented inline in the standard reconstruction framework of the scanner. Spatial and temporal L1 regularisation was performed for the frames  $\{\mathbf{x}_t\}_{t=1,\dots,T}$  for all time points  $T$  as similarly in previous work [25]:

$$\{\mathbf{x}_t\}_{t=1,\dots,T} = \operatorname{argmin}_{\{\mathbf{x}_t\}} \sum_{t=1}^T (\|\mathbf{A}_t \mathbf{x}_t - \mathbf{y}_t\|_2^2 + \lambda_\sigma \|\mathbf{W}_\sigma \mathbf{x}_t\|_1) + \lambda_\tau \|\mathbf{W}_\tau \{\mathbf{x}_1^\top, \dots, \mathbf{x}_T^\top\}^\top\|_1, \quad (1)$$

$\mathbf{A}_t$  is the system matrix for time  $t$  consisting of the corresponding sampling pattern, Fourier transform, and coil sensitivity maps for the local receiver coil elements. The measured data for time  $t$  is denoted by  $\mathbf{y}_t$ ,  $\lambda_\sigma$  and  $\lambda_\tau$  are the spatial and temporal regularisation parameters respectively.  $\mathbf{W}_\sigma$  and  $\mathbf{W}_\tau$  are the corresponding spatial and temporal Wavelet transforms respectively. Equation 1 is solved using Fast Iterative Shrinkage Thresholding Algorithm (FISTA) optimisation [26] alternating a gradient descent step for the quadratic terms and the evaluation of the proximal operator of the L1 terms. The proximal step was computed using the memory-efficient algorithm [27], and a total of 40 iterations were used for each reconstruction.

To evaluate and optimise the weight of the spatio-temporal regularisation terms ( $\lambda_\sigma$  and  $\lambda_\tau$ ), seven different reconstructions were performed for each patient. The first four reconstructions evaluated the impact of increasing both spatial and temporal regularisation as follows with an approximate doubling of the regularisation factors in succession: SMS-iter 1 ( $\lambda_\sigma$  0.0005  $\lambda_\tau$  0.0005), SMS-iter 2 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.001), SMS-iter 3 ( $\lambda_\sigma$  0.0025  $\lambda_\tau$  0.0025), SMS-iter 4 ( $\lambda_\sigma$  0.005  $\lambda_\tau$  0.005). The subsequent three reconstructions evaluated the impact of using a greater weighting for temporal regularisation. As SMS-iter 2 was found superior among the first four reconstructions (as described in the result section), the spatial regularisation factor ( $\lambda_\sigma$ ) was kept

constant to 0.001, whilst the temporal regularisation factor ( $\lambda_T$ ) was almost doubled in succession: SMS-iter 5 ( $\lambda_\sigma$  0.001  $\lambda_T$  0.002), SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_T$  0.005), SMS-iter 7 ( $\lambda_\sigma$  0.001  $\lambda_T$  0.01). For comparison, 3 slice bSSFP and SMS data were reconstructed using standard TGRAPPA reconstruction [28].

#### Qualitative image assessment

Qualitative image analysis was undertaken in consensus by two experienced CMR readers (AC and TI) with more than 10 years' experience in CMR each using a standardised rating scale (Table 1). Overall diagnostic image quality, perceived SNR, 'sequence related' artifact and 'patient related' artifact were ranked for each perfusion dataset (Table 1). The CMR readers were blinded to the clinical details of the patients and to the method of reconstruction for SMS imaging. Images were presented to readers in randomised order.

#### Quantitative assessment

In order to provide quantitative metrics for image quality, contrast ratio (CR) and extent of dark rim artifact (DRA) were evaluated. For CR, regions of interest (ROI) from the mid ventricular slice of the perfusion sequence were taken, in order to avoid partial volume effects of sampling at the basal or apical slice. ROIs for the myocardium were obtained with manual contouring of endocardium and epicardium and of the LV blood pool with careful exclusion of papillary muscles. CR was calculated as the ratio of peak LV blood pool SI : peak myocardial SI per slice. The extent of DRA was defined as percentage of the circumferential DRA of the total endocardium.

## Statistical Analysis

All statistical analyses were performed using SPSS Statistics 23 (IBM, Armonk, NY, USA). Results are expressed as mean  $\pm$  standard deviation unless otherwise specified. Qualitative image quality, sequence related artifact and patient related artifact scores, and perceived SNR were compared between methods using the Wilcoxon signed ranks test for paired ordinal data. Mean CR scores were compared between methods using paired t tests, having checked the assumption of normally distributed differences. All statistical tests were two-tailed and  $p$ -values  $< 0.05$  were considered significant.

## **Results**

### Study population

All patient scans were completed successfully. The CMR examination was normal in 6 of the patients. 2 patients were found to have sustained previous myocardial infarction and had impaired left ventricular systolic function. Participant characteristics are shown in Table 2. Two patients had suboptimal breath holds during contrast administration, reflecting real-world clinical practice.

### Slice location and image reconstruction

Using the '3 of 5' approach for slice location, reliable basal, mid and apical slices were generated for all 3 slice bSSFP images as defined by established criteria for slice location [29]. The '6 of 10' approach generated reliable basal, mid and apical slices in 46/48 (96%) of all SMS slices. Of the remaining 2/48 slices, in two patients,

the basal LV slice included part of the LVOT. All 3 slice bSSFP data were reconstructed with TGRAPPA and SMS data were successfully reconstructed with TGRAPPA and the different iterative reconstruction parameters on the scanner platform. The iterative reconstruction of the SMS images took approximately 10 minutes on the scanner console.

#### Optimum SMS iterative reconstruction

As the weighting of spatial ( $\lambda_\sigma$ ) and temporal ( $\lambda_\tau$ ) regularisation were both sequentially increased (SMS-iter 1-4, see Table 3), there was a trend towards increased perceived SNR and CR. However, for high spatial and temporal regularisation (SMS-iter 3 and 4), higher sequence related artifact (due to increased frequency of respiratory ghosting and image blurring) was observed resulting in a reduction in overall image quality. Among these four SMS iterative reconstructions, SMS-iter 2 was used as a basis for further investigation of temporal regularisation. Overall, the optimal SMS iterative reconstruction method with the ranking for the best overall image quality ( $2.50 \pm 0.53$ ), least sequence related artifact ( $0.13 \pm 0.35$ ) and highest perceived SNR ( $2.25 \pm 0.46$ ) was SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) (Table 3). This apparent difference is well illustrated in one patient as shown in Video 1. Therefore, SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) was chosen as the optimal SMS iterative reconstruction method for subsequent comparisons between 3 slice bSSFP and SMS-TGRAPPA. A comparison of 3 slice bSSFP and the optimum SMS iterative reconstruction method [SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005)] for two patients is presented in Figure 1 and Figure 2.

In two patients with suboptimal breath holds, there was degradation in overall image quality and increased sequence related artifact with increased respiratory ghosting in the iterative reconstruction parameters with greater spatial regularisation ( $\lambda_\sigma$ ) and temporal regularisation ( $\lambda_\tau$ ), in particular for SMS-iter 4 ( $\lambda_\sigma$  0.0005  $\lambda_\tau$  0.0005). With SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005), the overall image quality remained diagnostic in these two patients, with good overall image quality and reduction in respiratory ghosting and high perceived SNR compared to SMS-iter 4 ( $\lambda_\sigma$  0.0005  $\lambda_\tau$  0.0005).

#### Qualitative image assessment

SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) had superior overall image quality ( $2.50 \pm 0.53$  vs  $1.50 \pm 0.53$ ,  $p=0.005$ ) (Figure 3A) and perceived SNR ( $2.25 \pm 0.46$  vs  $0.75 \pm 0.46$ ,  $p=0.010$ ) compared to the 3 slice bSSFP (Figure 3B). There was no significant difference in sequence related artifact with SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) compared to the 3 slice bSSFP ( $0.13 \pm 0.35$  vs  $2.50 \pm 3.55$ ,  $p=0.14$ ). Importantly, no banding artifact was observed in any of the SMS reconstructions over the myocardium. With SMS-TGRAPPA compared to 3 slice bSSFP, there were no significant differences in overall image quality, sequence related artifact or perceived SNR. SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) compared to a SMS-TGRAPPA reconstruction had better overall image quality ( $2.50 \pm 0.53$  vs  $1.13 \pm 0.64$ ,  $p=0.015$ ) and better perceived SNR ( $2.25 \pm 0.46$  vs  $0.63 \pm 0.52$ ,  $p=0.009$ ). Interestingly, SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) was associated with a reduction in sequence related artifacts compared to SMS-TGRAPPA ( $0.13 \pm 0.35$  vs  $1.38 \pm 1.79$ ,  $p=0.043$ ) which was due to a reduction in respiratory ghosting.

Overall, there were no significant differences for patient related artifact between 3 slice bSSFP, SMS-TGRAPPA or SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005).

#### Quantitative image assessment

SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) had similar CR compared to 3 slice bSSFP ( $3.62 \pm 0.39$  vs  $3.66 \pm 0.65$ ,  $p=0.89$ ) (Figure 3C). There was no significant difference in CR between SMS-TGRAPPA and 3 slice bSSFP ( $3.30 \pm 0.34$  vs  $3.66 \pm 0.65$ ,  $p=0.20$ ). CR with SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) was higher than SMS-TGRAPPA ( $3.62 \pm 0.39$  vs  $3.30 \pm 0.34$ ,  $p=0.013$ ). There were no significant differences in % DRA between 3 slice bSSFP and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) ( $5.25 \pm 6.56$  vs  $4.25 \pm 4.30$ ,  $p=0.64$ ), 3 slice bSSFP and SMS-TGRAPPA ( $5.25 \pm 6.56$  vs  $4.37 \pm 4.43$ ,  $p=0.59$ ) and SMS-TGRAPPA and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) ( $4.37 \pm 4.43$  vs  $4.25 \pm 4.30$ ,  $p = 0.92$ ) (Figure 3D).



## Discussion

We demonstrated the clinical feasibility of SMS contrast enhanced first-pass myocardial perfusion imaging with bSSFP, GC-LOLA, and iterative reconstruction at 1.5 Tesla, which is a prerequisite prior to clinical evaluation in patients with suspected CAD for potential future clinical application. Doubled spatial coverage was achieved with SMS compared to a bSSFP approach with preserved spatial resolution. The employed iterative reconstruction technique of SMS data led to superior overall image quality, superior perceived SNR and similar CR compared to the 3 slice bSSFP. No banding artifacts were observed in any of the SMS perfusion images. Finally, a comprehensive image rating scale is proposed for application to development of myocardial perfusion sequences that may have utility to decipher optimal sequences and reconstruction methods.

Whole heart coverage for myocardial perfusion imaging is desirable as a strong correlation between CMR and nuclear perfusion studies has been demonstrated for the assessment of ischaemic burden [30], which is an important marker of prognosis [31]. High resolution myocardial perfusion imaging has been shown to improve detection of significant CAD through better detection of subendocardial ischaemia and less DRA [32]. In addition, high-resolution stress perfusion CMR allows for evaluation of transmural perfusion gradients to detect haemodynamically significant CAD [33, 34]. Currently, whole heart coverage can be achieved using 3D acquisition techniques which are associated with reduced in-plane spatial resolution while high resolution is achieved using multi-slice 2D acquisition protocols with limited spatial coverage (3-4 slices).

1 It is plausible that combining high resolution myocardial perfusion imaging with  
2 greater spatial coverage is advantageous for greater diagnostic accuracy and may  
3 provide a more accurate assessment of ischaemic burden. SMS may achieve the  
4 potential synergy of greater spatial coverage and high in-plane spatial resolution and  
5 this requires formal clinical evaluation in patients with CAD in a future clinical study.

6  
7 The feasibility of CAIPIRINHA perfusion CMR has previously been demonstrated in  
8 a small cohort of healthy volunteers using GRE readout [35-37] and combined with  
9 iterative regularised reconstruction with radial acquisition [38]. Balanced steady state  
10 free precession (bSSFP) pulse sequences for myocardial perfusion imaging are  
11 attractive due to better SNR and contrast-to-noise ratio (CNR) compared with spoiled  
12 gradient echo (GRE) readout [39]. Previous work with CAIPIRINHA bSSFP  
13 myocardial perfusion imaging without GC-LOLA demonstrated an increased  
14 sensitivity of SMS to banding artifacts at 1.5T [15]. In the present study, we  
15 combined CAIPIRINHA bSSFP with GC-LOLA and iterative reconstruction and  
16 observed no banding artifacts over the myocardium. The findings in the current study  
17 confirm previous work that GC-LOLA reduces SMS related banding artifacts [16].  
18 Similar acquisition times were achieved compared to 3 slice bSSFP, which is  
19 important in clinical practice in order to avoid cardiac motion, particularly in stress  
20 perfusion imaging with greater heart rates.

21  
22 Alternatively, SMS with bSSFP can be performed using blipped-CAIPI encoding  
23 where additional slice-gradient blips are employed to generate k-space phase  
24 modulation [40-42]. Although the employed approach has the potential to offer

reduced sensitivity to eddy currents when compared to blipped-CAIPI encoding [16], future studies are warranted to compare both approaches.

There was lower perceived SNR for SMS-TGRAPPA compared to a 3 slice bSSFP approach. We consider the reduction in perceived SNR for SMS-TGRAPPA to be related to additional g-factor noise amplification [13, 43], which increases with the overall acceleration factor. However, the potential loss in SNR was recovered with the optimal iterative reconstruction parameters and resulted in an improved overall image quality and perceived SNR compared to a bSSFP approach and SMS-TGRAPPA.

Sequence related artifacts increased with greater spatial and temporal regularisation and reduced overall image quality. There was also a trend for increased respiratory ghosting in iterative reconstruction of SMS data with greater spatial regularisation ( $\lambda_\sigma$ ) which indicates that such reconstruction parameters may not be suitable in patients with poor breath holds. In addition, in two patients with poor breath holds, increased artifacts were observed with a reduction in image quality in reconstructions with high spatial and temporal regularisation. However, using the optimal iterative reconstruction, there was a reduction in sequence related artifact and diagnostic image quality was achieved in all patients including two patients with poor breath holding.

Using the rankings obtained from the rating scheme presented in Table 1, SMS-iter 6 ( $\lambda_\sigma$  0.001  $\lambda_\tau$  0.005) was identified as the optimal imaging reconstruction parameters for SMS imaging, selected by the best overall image quality and perceived SNR and lowest frequency of sequence related artifact. Using this detailed rating scheme

1 allowed us to carefully decipher the optimal reconstruction parameters for the range  
2 of iterative reconstruction from 56 imaging datasets.

3  
4 In this study, we undertook myocardial perfusion imaging at rest without the  
5 administration of intravenous vasodilatory stress agents. In one patient with an  
6 ischaemic cardiomyopathy, and subendocardial myocardial infarction, rest perfusion  
7 imaging correctly delineated perfusion defects with areas of subendocardial scar on  
8 late gadolinium enhancement imaging (Figure 4). This indicates a signal for potential  
9 utility for application for ischaemia testing with SMS.

10 Vasodilatory stress increases myocardial blood flow (MBF) up to five-fold in healthy  
11 individuals, and in turn leads to a significant increase in signal intensity. This  
12 magnitude of signal intensity increase is not seen with rest perfusion imaging and  
13 this may reflect the overall lower global image quality observed in the 3 slice bSSFP  
14 and SMS TGRAPPA. Nevertheless, by using rest perfusion imaging alone, with a  
15 lower resting MBF and subsequent lower signal intensity, the standards and  
16 benchmark for comparison are higher.

17 There are various confounding physiological factors when comparing repeated  
18 stress perfusion imaging due to absolute changes in haemodynamic responses [44],  
19 MBF, signal intensity and therefore image quality. Hence, for the purpose of this  
20 study, which serves to demonstrate the feasibility of combining SMS, bSSFP, GC-  
21 LOLA and iterative reconstruction, rest perfusion imaging only was used. Therefore,  
22 this study serves as an important step for the methods development prior to a clinical  
23 validation study in patients with stress perfusion imaging.

The prolonged computation times for iterative reconstruction for SMS data may pose a barrier to implementation into routine clinical practice. While iterative reconstruction can significantly improve CMR image quality, such an approach is computationally intensive compared to standard reconstruction [43]. Techniques to reduce reconstruction time such as by use of a graphics processing unit are feasible, have been applied to MRI data with iterative reconstruction [44], with substantial increase of reconstruction speed [45]. Hence, rapid processing of SMS data may be feasible with dedicated hardware for reconstruction methods and thereby facilitate implementation into routine clinical care.

## **Future work**

A slice acceleration factor of 2 was used for this study which resulted in acquisition of 6 slices. For true whole heart coverage with contiguous slice coverage and in particular the true apical cap, a slice acceleration factor of 3 or 4 would be required, but this would require a trade off against any potential g-factor noise amplification. Myocardial perfusion imaging at 3 Tesla is highly desirable with benefit of increased SNR and CNR, which can be traded off with higher acceleration with parallel imaging which comes with an SNR penalty [45]. Hence, SMS bSSFP GC-LOLA with iterative reconstruction could be evaluated at 3 Tesla field strength and is well suited for greater slice acceleration. However, increasing field strength may also lead to an increase in field inhomogeneities and in turn lead to greater banding artifact. Careful shimming and selection of the optimal frequency from a frequency scout can be used to minimise off-resonance artifacts [46].

1 In this study we demonstrated the feasibility of SMS in patients with rest perfusion  
2 imaging only in order to ascertain diagnostic image quality and determine the optimal  
3 reconstruction parameters for SMS imaging as a methods development study. In  
4 order to validate the clinical application of this technique, vasodilator stress in a large  
5 cohort of patients with suspected CAD would be required. This larger cohort would  
6 need to reflect the wide distribution of disease of CAD (single vessel, two vessel and  
7 multivessel) in addition to the variation of clinical factors (arrhythmias, poor breath  
8 holders and obese patients) encountered in clinical practice. This current study now  
9 paves the way for such a clinical study in a group of patients with correlation of  
10 ischaemia related perfusion with invasive coronary angiographic fractional flow  
11 reserve data and/or PET.

## 13 **Limitations**

14 The sample size for this study is modest, but the purpose of this study was to  
15 determine the feasibility of SMS bSSFP with GC-LOLA and iterative reconstruction  
16 and to compare to a standard sequence used for routine clinical practice.  
17 Undertaking SNR and CNR measurements with parallel imaging are challenging and  
18 the added complexity in this study is that iterative reconstruction inherently  
19 thresholds and shrinks noise inhomogeneously across the field of view [47]. Studies  
20 with compressed sensing have avoided reporting absolute SNR measurements [48]  
21 and reported visual perception of SNR on a four point scale [47]. In this study, we  
22 also reported perceived SNR with detailed explanation of each parameter and  
23 calibration between observers. In addition, we reported CR to provides a metric for

quantitative image quality, which is not absolute SNR or CNR, but provides a meaningful ratio for image quality from dynamic perfusion images.

A minimum duration of 10 minutes is recommended between repeat contrast myocardial perfusion imaging [49], although contrast retention is often observed with longer periods. This study protocol used a minimum washout period of 15 minutes between each contrast administration. The time period for washout of contrast may have influenced the baseline signal intensity for the second rest perfusion study undertaken, although by alternating the sequences in each successive patient, we attempted to counterbalance the overall effect in this cohort of patients.

While 7 different combinations of weighting for spatial and temporal regularisation were employed, further combinations could have explored the effect of greater spatial and / or temporal regularisation. However, by using a step wise range of permutations for regularisation, we attempted to encompass a wide range of possible reconstructions parameters. In addition, greater regularisation may artificially over-smooth the images, with loss of important spatial and temporal data for dynamic perfusion imaging and hence we chose to limit the extent of regularisation.

The number of slices for comparison between 3 slice bSSFP and 6-slice SMS were not equal, and this may have influenced the comparability of the ratings presented in table 1. In any case, such an effect would bias against the new proposed technique of SMS as if an artifact was observed in 1/3 bSSFP images, this would score the same as an artifact in 1 or 2 SMS slices. Image ratings were performed with all slices together for each perfusion sequence rather than singles slices in isolation in order to allow global assessment of image quality and artifacts. Ratings could have been

1 taken individually for each slice, although in clinical practice, dynamic perfusion  
2 images are interpreted collectively rather than on an individual slice, and hence we  
3 chose to rate all perfusion slices for each dataset collectively.

## 4 **Conclusion**

6 Contrast enhanced myocardial perfusion imaging using SMS bSSFP with GC-LOLA  
7 and iterative reconstruction is feasible and provides improved image quality, doubled  
8 spatial coverage and identical in-plane spatial resolution compared to a 3 slice  
9 bSSFP approach. This technique may represent a route to achieve high resolution  
10 3D whole heart coverage for improved diagnostic accuracy, identification of  
11 subendocardial ischaemia and assessment of ischaemic burden in patients with  
12 suspected CAD. A clinical validation study in patients with CAD is now warranted.



## 1 **List of Abbreviations**

- 2 bSSFP: balanced Steady State Free Precession
- 3 CAD: Coronary Artery Disease
- 4 CAIPIRINHA: Controlled Aliasing in Parallel Imaging Results in Higher Acceleration
- 5 CMR: Cardiac Magnetic Resonance
- 6 CNR: Contrast to Noise Ratio
- 7 DRA: Dark Rim Artifact
- 8 ECG: Electrocardiogram
- 9 EDV: End Diastolic Volume
- 10 FISTA: Fast Iterative Shrinkage Thresholding Algorithm
- 11 GC-LOLA: Gradient Controlled Local Larmor Adjustment
- 12 GRAPPA: Generalized Autocalibrating Partial Parallel Acquisition
- 13 GRE: Gradient Recalled Echo
- 14 LV: Left Ventricle
- 15 LVEF: Left ventricular ejection fraction
- 16 LVOT: Left Ventricular Outflow Tract
- 17 MRI: Magnetic Resonance Imaging
- 18 MI: myocardial infarction
- 19 PET: Positron Emission Tomography
- 20 POMP: Phase Offset Multiplanar
- 21 RF: Radiofrequency
- 22 RVEF: Right ventricular ejection fraction
- 23 SENSE: Sensitivity Encoding
- 24 SI: Signal Intensity

- 1 SMS: Simultaneous Multi Slice
- 2 SMS-iter: Simultaneous Multi Slice with Iterative reconstruction
- 3 SNR: Signal to Noise Ratio
- 4 TI: inversion time
- 5 TE: echo time
- 6 TR: repetition time
- 7 TGRAPPA: Temporal GRAPPA

## 1   **Declarations**

## 2   **Ethics approval and consent to participate**

3   The study was approved by the National Research Ethics Service (15/NS/0030 for the  
4   patient study). Inform consent was obtained from all participants.

## 6   **Consent for publication**

7   Consent for publication was obtained from all participants in the study.

## 9   **Availability of data and material**

10   The datasets used and/or analyzed during the current study are available from the  
11   corresponding author on reasonable request.

## 13   **Competing interests**

14   The authors declare that they have no competing interests

## 16   **Funding**

17   The authors acknowledge financial support from the Department of Health through  
18   the National Institute for Health Research (NIHR) comprehensive Biomedical  
19   Research Centre award to Guy's & St Thomas' NHS Foundation Trust in partnership  
20   with King's College London and King's College Hospital NHS Foundation Trust and  
21   by the NIHR Healthcare Technology Co-operative for Cardiovascular Disease at  
22   Guy's and St Thomas' NHS Foundation Trust. This work was supported by the  
23   Wellcome/EPSRC Centre for Medical Engineering [WT 203148/Z/16/Z] and the  
24   EPSRC grant [EP/R010935/1]. MSN was funded by the UK Medical Research

Council under grant number MR/P01979X/1. SP was funded by a British Heart Foundation Chair under grant number CH/16/2/32089. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR, the DoH, EPSRC, MRC or the Wellcome Trust.

#### **Authors' contributions**

SR and MSN designed the study protocol, acquired and analysed the data and drafted the manuscript. AC, TI and RN assisted with study design and undertook data analysis and interpretation and critically revised the manuscript. FR undertook the statistical analysis and provided expert statistical support. PS, DS, SP, RR, CF and MS assisted with study design, interpretation of data and critically revised the manuscript. All authors read and approved the final manuscript.

#### **Acknowledgements:**

The authors would like to thank the radiographers and administration team at King's College London and the Guy's and St Thomas' NHS Hospital Cardiovascular MRI Service for their cooperation and assistance during the imaging and administration processes.

## References

1. Fihn SD, Gardin JM, Abrams J, Berra K, Blankenship JC, Dallas AP, Douglas PS, Foody JM, Gerber TC, Hinderliter AL *et al*: **2012 ACCF/AHA/ACP/AATS/PCNA/SCAI/STS Guideline for the diagnosis and management of patients with stable ischemic heart disease: a report of the American College of Cardiology Foundation/American Heart Association Task Force on Practice Guidelines, and the American College of Physicians, American Association for Thoracic Surgery, Preventive Cardiovascular Nurses Association, Society for Cardiovascular Angiography and Interventions, and Society of Thoracic Surgeons.** *J Am Coll Cardiol* 2012, **60**(24):e44.
2. Montalescot G, Sechtem U, Achenbach S, Andreotti F, Arden C, Budaj A, Bugiardini R, Crea F, Cuisset T, Di Mario C *et al*: **2013 ESC guidelines on the management of stable coronary artery disease: the Task Force on the management of stable coronary artery disease of the European Society of Cardiology.** *Eur Heart J* 2013, **34**(38):2949-3003.
3. Jaarsma C, Leiner T, Bekkers SC, Crijns HJ, Wildberger JE, Nagel E, Nelemans PJ, Schalla S: **Diagnostic performance of noninvasive myocardial perfusion imaging using single-photon emission computed tomography, cardiac magnetic resonance, and positron emission tomography imaging for the detection of obstructive coronary artery disease: a meta-analysis.** *J Am Coll Cardiol* 2012, **59**(19):1719-1728.
4. Kramer CM, Barkhausen J, Flamm SD, Kim RJ, Nagel E, Society for Cardiovascular Magnetic Resonance Board of Trustees Task Force on Standardized P: **Standardized cardiovascular magnetic resonance (CMR) protocols 2013 update.** *J Cardiovasc Magn Reson* 2013, **15**(1):91.
5. Motwani M, Jogiya R, Kozerke S, Greenwood JP, Plein S: **Advanced cardiovascular magnetic resonance myocardial perfusion imaging: high-spatial resolution versus 3-dimensional whole-heart coverage.** *Circ Cardiovasc Imaging* 2013, **6**(2):339-348.
6. Fair MJ, Gatehouse PD, DiBella EV, Firmin DN: **A review of 3D first-pass, whole-heart, myocardial perfusion cardiovascular magnetic resonance.** *J Cardiovasc Magn Reson* 2015, **17**:68.
7. Souza SP, Szumowski J, Dumoulin CL, Plewes DP, Glover G: **SIMA: simultaneous multislice acquisition of MR images by Hadamard-encoded excitation.** *J Comput Assist Tomogr* 1988, **12**(6):1026-1030.
8. Muller S: **Simultaneous multislice imaging (SIMUSIM) for improved cardiac imaging.** *Magn Reson Med* 1989, **10**(1):145-155.
9. Glover GH: **Phase-offset multiplanar (POMP) volume imaging: a new technique.** *J Magn Reson Imaging* 1991, **1**(4):457-461.
10. Larkman DJ, Hajnal JV, Herlihy AH, Coutts GA, Young IR, Ehnholm G: **Use of multicoil arrays for separation of signal from multiple slices simultaneously excited.** *J Magn Reson Imaging* 2001, **13**(2):313-317.
11. Breuer FA, Blaimer M, Heidemann RM, Mueller MF, Griswold MA, Jakob PM: **Controlled aliasing in parallel imaging results in higher acceleration (CAIPIRINHA) for multi-slice imaging.** *Magn Reson Med* 2005, **53**(3):684-691.

12. Griswold MA, Jakob PM, Heidemann RM, Nittka M, Jellus V, Wang J, Kiefer B, Haase A: **Generalized autocalibrating partially parallel acquisitions (GRAPPA)**. *Magn Reson Med* 2002, **47**(6):1202-1210.
13. Pruessmann KP, Weiger M, Scheidegger MB, Boesiger P: **SENSE: sensitivity encoding for fast MRI**. *Magn Reson Med* 1999, **42**(5):952-962.
14. Sodickson DK, Manning WJ: **Simultaneous acquisition of spatial harmonics (SMASH): fast imaging with radiofrequency coil arrays**. *Magn Reson Med* 1997, **38**(4):591-603.
15. Stab D, Ritter CO, Breuer FA, Weng AM, Hahn D, Kostler H: **CAIPIRINHA accelerated SSFP imaging**. *Magn Reson Med* 2011, **65**(1):157-164.
16. Stab D, Speier P: **Gradient-controlled local Larmor adjustment (GC-LOLA) for simultaneous multislice bSSFP imaging with improved banding behavior**. *Magn Reson Med* 2018. [doi: 10.1002/mrm.27356]
17. Lin FH, Kwong KK, Belliveau JW, Wald LL: **Parallel imaging reconstruction using automatic regularization**. *Magn Reson Med* 2004, **51**(3):559-567.
18. Peng Q, Jing L, Bida Z, Jianmin W, X. SG: **An improved iterative SENSE reconstruction method**. *Concepts in Magnetic Resonance Part B: Magnetic Resonance Engineering* 2007, **31B**(1):44-50.
19. Liang D, Liu B, Wang J, Ying L: **Accelerating SENSE using compressed sensing**. *Magn Reson Med* 2009, **62**(6):1574-1584.
20. Ramani S, Fessler JA: **Parallel MR image reconstruction using augmented Lagrangian methods**. *IEEE Trans Med Imaging* 2011, **30**(3):694-706.
21. Lustig M, Donoho D, Pauly JM: **Sparse MRI: The application of compressed sensing for rapid MR imaging**. *Magn Reson Med* 2007, **58**(6):1182-1195.
22. Ishida M, Schuster A, Morton G, Chiribiri A, Hussain S, Paul M, Merkle N, Steen H, Lossnitzer D, Schnackenburg B *et al*: **Development of a universal dual-bolus injection scheme for the quantitative assessment of myocardial perfusion cardiovascular magnetic resonance**. *J Cardiovasc Magn Reson* 2011, **13**:28.
23. Plein S, Greenwood J, Ridgway JP: **Cardiovascular MR Manual**. 2015.
24. Beck A, Teboulle M: **A Fast Iterative Shrinkage-Thresholding Algorithm for Linear Inverse Problems**. *SIAM Journal on Imaging Sciences* 2009, **2**(1):183-202.
25. Wetzl J, Forman C, Wintersperger BJ, D'Errico L, Schmidt M, Mailhe B, Maier A, Stalder AF: **High-resolution dynamic CE-MRA of the thorax enabled by iterative TWIST reconstruction**. *Magn Reson Med* 2017, **77**(2):833-840.
26. Liu J, Lefebvre A, Zenge MO, Schmidt M, Mueller E, Nadar MS: **2D bSSFP real-time cardiac CINE-MRI: compressed sensing featuring weighted redundant Haar Wavelet regularization in space and time**. *J Cardiovasc Magn Reson* 2013, **15**(Suppl 1):P49-P49.
27. Chambolle A, Pock T: **A First-Order Primal-Dual Algorithm for Convex Problems with Applications to Imaging**. *Journal of Mathematical Imaging and Vision* 2011, **40**(1):120-145.
28. Breuer FA, Kellman P, Griswold MA, Jakob PM: **Dynamic autocalibrated parallel imaging using temporal GRAPPA (TGRAPPA)**. *Magn Reson Med* 2005, **53**(4):981-985.

29. Cerqueira MD, Weissman NJ, Dilsizian V, Jacobs AK, Kaul S, Laskey WK, Pennell DJ, Rumberger JA, Ryan T, Verani MS: **Standardized Myocardial Segmentation and Nomenclature for Tomographic Imaging of the Heart. A Statement for Healthcare Professionals From the Cardiac Imaging Committee of the Council on Clinical Cardiology of the American Heart Association** 2002, **105**(4):539-542.
30. Jogiya R, Morton G, De Silva K, Reyes E, Hachamovitch R, Kozerke S, Nagel E, Underwood SR, Plein S: **Ischemic burden by 3-dimensional myocardial perfusion cardiovascular magnetic resonance: comparison with myocardial perfusion scintigraphy.** *Circ Cardiovasc Imaging* 2014, **7**(4):647-654.
31. Shaw LJ, Berman DS, Maron DJ, Mancini GB, Hayes SW, Hartigan PM, Weintraub WS, O'Rourke RA, Dada M, Spertus JA *et al*: **Optimal medical therapy with or without percutaneous coronary intervention to reduce ischemic burden: results from the Clinical Outcomes Utilizing Revascularization and Aggressive Drug Evaluation (COURAGE) trial nuclear substudy.** *Circulation* 2008, **117**(10):1283-1291.
32. Motwani M, Maredia N, Fairbairn TA, Kozerke S, Radjenovic A, Greenwood JP, Plein S: **High-resolution versus standard-resolution cardiovascular MR myocardial perfusion imaging for the detection of coronary artery disease.** *Circ Cardiovasc Imaging* 2012, **5**(3):306-313.
33. Hautvast GL, Chiribiri A, Lockie T, Breeuwer M, Nagel E, Plein S: **Quantitative analysis of transmural gradients in myocardial perfusion magnetic resonance images.** *Magn Reson Med* 2011, **66**(5):1477-1487.
34. Chiribiri A, Hautvast GL, Lockie T, Schuster A, Bigalke B, Olivotti L, Redwood SR, Breeuwer M, Plein S, Nagel E: **Assessment of coronary artery stenosis severity and location: quantitative analysis of transmural perfusion gradients by high-resolution MRI versus FFR.** *JACC Cardiovasc Imaging* 2013, **6**(5):600-609.
35. Stab D, Wech T, Breuer FA, Weng AM, Ritter CO, Hahn D, Kostler H: **High resolution myocardial first-pass perfusion imaging with extended anatomic coverage.** *J Magn Reson Imaging* 2014, **39**(6):1575-1587.
36. Weingärtner S, Moeller S, Akcakaya M: **Feasibility of Ultra-high Simultaneous Multi-slice and In-plane Accelerations for Cardiac MRI Using Outer Volume Suppression and Leakage-Blocking Reconstruction.** In: *International Society for Magnetic Resonance in Medicine*. Paris, France; 2018.
37. Yang Y, Houten MV, Norton P, Hagspiel K, Kramer C, Mugler J, Salerno M: **Spiral simultaneous multi-slice first-pass myocardial perfusion imaging.** In: *International Society for Magnetic Resonance in Medicine*. Honolulu, USA; 2017.
38. Wang H, Adluru G, Chen L, Kholmovski EG, Bangerter NK, DiBella EV: **Radial simultaneous multi-slice CAIPI for ungated myocardial perfusion.** *Magn Reson Imaging* 2016, **34**(9):1329-1336.
39. Kellman P, Arai AE: **Imaging sequences for first pass perfusion --a review.** *J Cardiovasc Magn Reson* 2007, **9**(3):525-537.

40. Price AN, Cordero-Grande L, Malik SJ, JV H: **Accelerated cine imaging of the heart using blipped multiband SSFP**. In: *Proceedings of the 25th Annual Meeting of ISMRM*. Honolulu, HI; 2017: 0631.
41. Setsompop K, Gagoski BA, Polimeni JR, Witzel T, Wedeen VJ, Wald LL: **Blipped-Controlled Aliasing in Parallel Imaging (blipped-CAIPI) for simultaneous multi-slice EPI with reduced g-factor penalty**. *Magn Reson Med* 2012, **67**(5):1210-1224.
42. Ye H, Ma D, Jiang Y, Cauley SF, Du Y, Wald LL, Griswold MA, Setsompop K: **Accelerating magnetic resonance fingerprinting (MRF) using t-blipped simultaneous multislice (SMS) acquisition**. *Magn Reson Med* 2016, **75**(5):2078-2085.
43. Breuer FA, Kannengiesser SA, Blaimer M, Seiberlich N, Jakob PM, Griswold MA: **General formulation for quantitative G-factor calculation in GRAPPA reconstructions**. *Magn Reson Med* 2009, **62**(3):739-746.
44. Schelbert HR: **Anatomy and physiology of coronary blood flow**. *J Nucl Cardiol* 2010, **17**(4):545-554.
45. Oshinski JN, Delfino JG, Sharma P, Gharib AM, Pettigrew RI: **Cardiovascular magnetic resonance at 3.0 T: current state of the art**. *J Cardiovasc Magn Reson* 2010, **12**:55.
46. Wansapura J, Fleck R, Crotty E, Gottliebson W: **Frequency scouting for cardiac imaging with SSFP at 3 Tesla**. *Pediatr Radiol* 2006, **36**(10):1082-1085.
47. Akçakaya M, Basha TA, Chan RH, Manning WJ, Nezafat R: **Accelerated Isotropic Sub-Millimeter Whole-Heart Coronary MRI: Compressed Sensing versus Parallel Imaging**. *Magn Reson Med* 2014, **71**(2):815-822.
48. Vasanawala SS, Alley MT, Hargreaves BA, Barth RA, Pauly JM, Lustig M: **Improved Pediatric MR Imaging with Compressed Sensing**. *Radiology* 2010, **256**(2):607-616.
49. Jerosch-Herold M: **Quantification of myocardial perfusion by cardiovascular magnetic resonance**. *J Cardiovasc Magn Reson* 2010, **12**:57.



Video 1. Dynamic perfusion images acquired from a mid-ventricular slice following contrast administration for a patient with a breath hold at peak contrast administration. The different reconstructions with simultaneous multi slice (SMS) with TGRAPPA and iterative reconstruction are presented. SMS-TGRAPPA was associated with poor SNR. SMS-iter 1 ( $\lambda_{\sigma}$  0.0005  $\lambda_{\tau}$  0.0005), with the least weighting for spatio-temporal regularisation had poor perceived SNR although moderate overall image quality. As the weighting of combined spatio-temporal regularisation increased, as with SMR-iter 4 ( $\lambda_{\sigma}$  0.005  $\lambda_{\tau}$  0.005), despite an improved SNR, there was greater sequence related artefact (particularly respiratory ghosting and image blurring) with a reduction overall image quality. The optimum SMS reconstruction SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) was determined to have the most favourable image quality, with the highest overall image quality, least sequence related artifacts and high perceived SNR.

Figure 1. Dynamic perfusion series following contrast administration using 3 slice bSSFP and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) in patient 1. Top to bottom: base to apex. Left to right, baseline images, contrast transit through right ventricle, left ventricular blood pool, peak myocardial and washout. SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) had better subjective image quality compared to 3 slice bSSFP.

Figure 2. Dynamic perfusion series following contrast administration using 3 slice bSSFP and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) in patient 2. Top to bottom: base to apex. Left to right, baseline images, contrast transit through right ventricle, left ventricular blood pool, peak myocardial and washout. SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) had better subjective image quality compared to 3 slice bSSFP.

Figure 3. Comparison of 3 slice bSSFP, SMS-TGRAPPA and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) in 8 patients. (A) Overall diagnostic image quality. Scores for image quality range from 0 to 3 (0= poor image quality and non-diagnostic, 1= major artifact present but not limiting diagnosis, 2= minor artifact present but not limiting diagnosis, 3= excellent). Overall image quality was significantly higher with SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) compared to SMS-TGRAPPA and 3 slice bSSFP. (B) Perceived Signal to Noise Ratio (SNR) with 3 slice bSSFP, SMS-TGRAPPA and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005). Scores for perceived SNR from 0 to 3 (0= very poor SNR non-diagnostic image quality, 1= major noise present but not limiting diagnosis, 2= minor noise present but not limiting diagnosis, 3= high SNR with excellent image quality). Perceived SNR was significantly higher with SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) compared to SMS-TGRAPPA and 3 slice bSSFP. (C) Contrast ratio (ratio peak blood

pool signal intensity : peak myocardial signal intensity). There was no significant difference in Contrast Ratio between SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) and 3 slice bSSFP. (D) Dark rim artifact (mean and standard deviation): There was no significant difference in the percentage of dark rim artifact between 3 slice bSSFP and SMS-TGRAPPA or SMS-TGRAPPA and SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005).

Figure 4. Peak myocardial perfusion signal intensity images and late gadolinium enhancement (LGE) images of a patient with subendocardial myocardial infarction. Top panel: peak myocardial dynamic frame for SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005); bottom panel: LGE imaging following contrast administration. The rest perfusion defects (black arrows) matched with the areas of subendocardial scar (white arrows).

1 Table 1. Four categories for qualitative image quality assessment. (A) overall  
2 diagnostic image quality (range 0-3), (B) sequence-related artifact (7 criteria range 0-  
3 3, maximum total score 21), (C) patient related artifact (2 criteria with range 0-3, total  
4 maximum score 6) and (D) perceived Signal-to-Noise-Ratio (SNR) (range 0-3). (B)  
5 Score for sequence related artifact relates to total number of slices acquired (artifact  
6 present in 1, 2 or 3 slices would score 1, 2 and 3 respectively for 3 slice bSSFP;  
7 artifact present in 1-2, 3-4 or 5-6 slices would score 1, 2 and 3 respectively for 6-slice  
8 SMS).

9

#### **Qualitative Criteria for perfusion Imaging**

	0	1	2	3	maximum score
<b>A. Overall Diagnostic Image Quality</b>					
	Poor image quality and non- diagnostic	major artifact present but not limiting diagnosis	Minor artifact present but not limiting diagnosis	Excellent	3
					3
<b>B. Sequence related artifact</b>					
Wrap around	none	1 slice	2 slices	3 slices	3
Respiratory ghosting	none	1 slice	2 slices	3 slices	3
Cardiac ghosting	none	1 slice	2 slices	3 slices	3
Image blurring	none	1 slice	2 slices	3 slices	3
Metallic artifact	none	1 slice	2 slices	3 slices	3
Banding artifact	none	1 slice	2 slices	3 slices	3
Cardiac Motion	none	1 slice	2 slices	3 slices	3
					21
<b>C. Patient / equipment related artifact</b>					
Breathing motion	none	Minor artifact present but not limiting diagnosis	major artifact present but not limiting diagnosis	Non- diagnostic	3
ECG mistriggers	None	1 mistriggers	2 mistriggers	> 2 mistriggers	3
					6

**D. Perceived Signal-to-Noise Ratio (SNR)**

	Very poor SNR non- diagnostic image quality	Minor noise present but not limiting diagnosis	major noise present but not limiting diagnosis	High SNR with excellent image quality	3
1					3

1 Table 2. Study participant characteristics for 8 patients. LVEF: Left ventricular  
2 ejection fraction; RVEF: Right ventricular ejection fraction; EDV: End Diastolic  
3 Volume; MI: myocardial infarction. Results are mean  $\pm$  standard deviation or number  
4 (%), as specified.

5

Age (years)	50 $\pm$ 22
Male: number (%)	6 (75%)
Body Mass Index (kg/m <sup>2</sup> )	25 $\pm$ 5
Previous MI: number (%)	2 (25%)
LVEF (%)	52 $\pm$ 16
Indexed LV EDV (ml/m <sup>2</sup> )	95 $\pm$ 27
RVEF (%)	57 $\pm$ 10
Indexed RV EDV (ml/m <sup>2</sup> )	80 $\pm$ 23
Scar present: number (%)	2 (25%)
Resting Heart Rate (beats/min)	65 $\pm$ 14

6

Table 3. Image quality assessment of images produced from iterative reconstruction of SMS images with different parameters, for 8 patients (see Table 1 for definition of rating scales).  $\lambda_{\sigma}$  indicates the degree of spatial regularisation, whilst  $\lambda_{\tau}$  indicates the extent of temporal regularisation. SMS-iter 6 ( $\lambda_{\sigma}$  0.001  $\lambda_{\tau}$  0.005) had the highest overall image quality and the lowest amount of sequence related artifact. Results are mean  $\pm$  standard deviation. Contrast ratio (CR) calculated as the ratio peak blood pool signal intensity : peak myocardial signal intensity.

Iterative reconstruction parameters	Overall Image Quality	Perceived SNR	Sequence Related Artifact	Patient Related Artifact	Contrast Ratio
SMS-iter 1 ( $\lambda_{\sigma}$ 0.0005 $\lambda_{\tau}$ 0.0005)	1.63 $\pm$ 0.74	1.63 $\pm$ 0.52	0.69 $\pm$ 1.10	0.89 $\pm$ 1.36	3.64 $\pm$ 0.37
SMS-iter 2 ( $\lambda_{\sigma}$ 0.001 $\lambda_{\tau}$ 0.001)	1.88 $\pm$ 0.35	1.63 $\pm$ 0.52	0.31 $\pm$ 0.59	0.89 $\pm$ 1.36	3.73 $\pm$ 0.44
SMS-iter 3 ( $\lambda_{\sigma}$ 0.0025 $\lambda_{\tau}$ 0.0025)	1.50 $\pm$ 0.53	2.00 $\pm$ 0.00	0.63 $\pm$ 1.27	1.00 $\pm$ 1.41	3.80 $\pm$ 0.47
SMS-iter 4 ( $\lambda_{\sigma}$ 0.005 $\lambda_{\tau}$ 0.005)	1.25 $\pm$ 0.46	1.88 $\pm$ 0.35	1.88 $\pm$ 1.62	1.00 $\pm$ 1.41	3.93 $\pm$ 0.51
SMS-iter 5 ( $\lambda_{\sigma}$ 0.001 $\lambda_{\tau}$ 0.0025)	2.13 $\pm$ 0.35	2.13 $\pm$ 0.35	0.25 $\pm$ 0.38	0.78 $\pm$ 1.40	3.63 $\pm$ 0.40
SMS-iter 6 ( $\lambda_{\sigma}$ 0.001 $\lambda_{\tau}$ 0.005)	2.50 $\pm$ 0.53	2.25 $\pm$ 0.46	0.13 $\pm$ 0.35	0.56 $\pm$ 1.01	3.62 $\pm$ 0.39
SMS-iter 7 ( $\lambda_{\sigma}$ 0.001 $\lambda_{\tau}$ 0.01)	2.00 $\pm$ 0.53	2.25 $\pm$ 0.46	0.56 $\pm$ 0.86	0.67 $\pm$ 1.32	3.64 $\pm$ 0.40